



CALIBRATION CERTIFICATE

CERT NO.: 4321
CUSTOMER: Sample
PO NUMBER: ABD12345
FILE NO.: SVC49876
CAL DATE: 08 September 2006
DUE DATE: 08 September 2007
CAL/TEST BY: E. Rennie

INSTRUMENT DETAILS

MANUFACTURER: MD Totco
DESCRIPTION: Wireline Weight Indicator System with 6" Gauge, 100ft Hose & 8"sq Tension Load Cell
RANGE: 0-4000 Lbs (FSD 0-700 PSI)
SERIAL NO.: Gauge 5110A, Cell 27DG703 **ID NO.:** ABC123456

TEST EQUIPMENT DETAILS

MANUFACTURER: DH Budenberg
DESCRIPTION: Model H540 Series Hydraulic Electronic Calibration System
RANGE: 0-1200 Bar / 17400 PSI
SERIAL NO.: H540/176

The accuracy of the test unit is traceable to National Standards through UKAS Cert Nos 02799 & 02979

SAMPLE TEST RESULTS

APPLIED PRESSURE	RISING PRESSURE	FALLING PRESSURE
PSI	Lbs	Lbs
0	0	0
140	800	810
280	1600	1610
420	2400	2410
560	3200	3205
700	4005	

FALLING READINGS ARE TAKEN WHEN RETURNING FROM MAXIMUM SCALE

"DUE DATE" IS GUIDE DATE ONLY AND MAY VARY DEPENDANT ON USEAGE.

E. Rennie

SIGNATURE: _____

Midcontinent (Aberdeen) Limited