



Midcontinent (Aberdeen) Limited
 Blackness Road
 Altens Industrial Estate
 Aberdeen AB12 3LH
 Tel: +44(0) 1224 874906
 Fax: +44(0)1224 898386
 e-mail: service@midcontinent.co.uk

WORK ORDER / ESTIMATE FORM

W.O. NUMBER.: 4321
 DATE: 07 September 2006
 CUSTOMER: Sample
 CONTACT: Sample
 REFERENCE: MAS987654321
 PRIORITY: ASAP

INSTRUMENT DETAILS

MANUFACTURER: MD Totco
 DESCRIPTION: Wireline Weight Indicator System with 6" Gauge, 100ft Hose & 8"sq Tension Load Cell
 RANGE: 0-4000 Lbs (FSD 0-700 PSI)
 SERIAL NO.: Gauge 5110A, Cell 27DG703
 ID NO.: ABC123456

CUSTOMER INSTRUCTIONS
 Repair and recalibrate.

WORK REQUIRED
 CLEAN BEADBLAST REPAINT CHECK ACCURACY ADJUST SPAN ADJUST LINEARITY RESET ZERO

PARTS REQUIRED

PART NO.	DESCRIPTION	QTY	EACH	TOTAL
MIDF101	Red Hydraulic Fluid	2	£4.00	£8.00
MID2345	Case Safety Relief Valve Assy	1	£23.50	£23.50
MID1234	Case Screws	3	£0.40	£1.20

SAMPLE

DELIVERY
 1-2 days from your go ahead.

PARTS TOTAL: £32.70
 LABOUR TOTAL: £60.00
TOTAL ESTIMATE FOR REPAIR: £92.70

COMMENTS / ADDITIONAL INFORMATION
 Reassembled and ready for use

TEST RESULTS

APPLIED PRESSURE	RISING PRESSURE	FALLING PRESSURE
PSI	Lbs	Lbs
0	0	0
140	800	810
280	1600	1610
420	2400	2410
560	3200	3205
700	4005	

PREPARED BY
 E. Rennie R. Goodbrand K. Mowat 3rd Party

TEST INSTRUMENT
 Budenberg H540 Electronic Comparator
 Thurlby 30V 2A PL320

FOR ADDITIONAL INFORMATION OR TO CONFIRM GO-AHEAD, PLEASE CONTACT MIDCONTINENT QUOTING THE ABOVE WORK NUMBER.

CUSTOMER REPLY SECTION:

CONTACT	P.O. NUMBER	COMMENTS